

# **Bodyguards** Statement of Product Conformity

**BM Polyco Ltd  
Crown Road  
Enfield  
EN1 1TX  
Great Britain**

declare that the new Medical Device described hereafter

## **Bodyguards Black Nitrile GL897**

A non-sterile, powder-free, nitrile rubber disposable glove

is in conformity with National Standards EN455-1, EN455-2, EN455-3

is Class I and is self-certified according to the requirements of EC Directive 93/42/EEC

### **Product Information**

| Size | Small  | Medium | Large  | Extra Large |
|------|--------|--------|--------|-------------|
| Code | GL8971 | GL8972 | GL8973 | GL8975      |

**Glove Care:** Store below 25C away from direct sunlight

**Thickness:** 0.08mm (measured at palm)

**Freedom from holes:** AQL 1.5

**Physical properties:** 6N minimum

This product does not contain natural rubber latex protein

This product contains low levels of residual chemical accelerators, which may cause allergic reactions.

For further information and advice on suitability for protection against specific substances or applications please contact us.